

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090719\
 Data File : VU034363.D
 Acq On : 07 Sep 2019 04:32
 Operator : JC/SP
 Sample : K4725-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S13-0263

Quant Time: Sep 07 06:27:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	294798	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	441485	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	439082	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	267654	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	220043	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
35) Dibromofluoromethane	4.86	113	195110	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
50) Toluene-d8	7.55	98	745213	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	10.29	95	271699	45.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.20%	
Target Compounds						
16) Acetone	2.31	43	15523	6.725	ug/l	89
39) Methylcyclohexane	6.40	83	8800	1.719	ug/l	94
95) Naphthalene	13.74	128	19679	4.982	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090719\
Data File : VU034363.D
Acq On : 07 Sep 2019 04:32
Operator : JC/SP
Sample : K4725-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S13-0263

Quant Time: Sep 07 06:27:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

