


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090719\
 Data File : VU034345.D
 Acq On : 06 Sep 2019 21:13
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Title : SW846 8260
 Last Update : Fri Sep 06 05:10:13 2019



AutoFind: Scans 2860, 2861, 2862; Background Corrected with Scan 2847

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	19549	PASS
75	95	30	60	46.4	56746	PASS
95	95	100	100	100.0	122232	PASS
96	95	5	9	6.3	7712	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.9	106202	PASS
175	174	5	9	7.5	7982	PASS
176	174	95	101	97.2	103184	PASS
177	176	5	9	6.5	6700	PASS