

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090721\
 Data File : VU044592.D
 Acq On : 07 Sep 2021 18:39
 Operator : SY/MD
 Sample : M3647-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F3W14

Quant Time: Sep 08 02:38:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 08 02:31:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	109312	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	114302	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	57491	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	34842	3.941	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.800%
7) Chloroethane-d5	1.919	69	36251	4.718	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.571	65	15185	4.038	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.800%
20) 2-Butanone-d5	4.642	46	105382	43.100	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.200%
24) Chloroform-d	5.070	84	67235	4.301	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
26) 1,2-Dichloroethane-d4	5.710	65	38211	4.482	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	5.735	84	140088	4.352	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.000%
36) 1,2-Dichloropropane-d6	6.697	67	42770	4.154	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.000%
41) Toluene-d8	7.902	98	121632	4.262	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	8.185	79	14500	3.921	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.400%
46) 2-Hexanone-d5	8.645	63	94407	47.210	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.420%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	39210	4.521	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.400%
66) 1,2-Dichlorobenzene-d4	12.198	152	48531	4.996	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%
Target Compounds					Qvalue	
25) Chloroform	5.095	83	2407	0.138	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090721\
Data File : VU044592.D
Acq On : 07 Sep 2021 18:39
Operator : SY/MD
Sample : M3647-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F3W14

Quant Time: Sep 08 02:38:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 08 02:31:37 2021
Response via : Initial Calibration

