

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090722\
 Data File : VU050664.D
 Acq On : 07 Sep 2022 16:38
 Operator : SY/MD
 Sample : N4482-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW346

Quant Time: Sep 08 05:43:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 07 01:37:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	214761	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	223253	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	80373	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	63089	34.346	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	68.700%		
7) Chloroethane-d5	1.906	69	52790	42.498	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	85.000%		
11) 1,1-Dichloroethene-d2	2.562	63	96895	29.850	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	59.700%#		
21) 2-Butanone-d5	4.629	46	155053	115.374	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	115.370%		
24) Chloroform-d	5.060	84	184514	53.000	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	106.000%		
26) 1,2-Dichloroethane-d4	5.700	65	130080	61.145	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	122.300%		
32) Benzene-d6	5.726	84	348181	47.193	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.380%		
36) 1,2-Dichloropropane-d6	6.691	67	119627	49.386	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	98.780%		
41) Toluene-d8	7.896	98	281511	44.619	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	89.240%		
43) trans-1,3-Dichloroprop...	8.179	79	52357	52.240	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	104.480%		
47) 2-Hexanone-d5	8.636	63	56619	92.926	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	92.930%		
56) 1,1,2,2-Tetrachloroeth...	10.755	84	193181	52.701	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	105.400%		
66) 1,2-Dichlorobenzene-d4	12.192	152	101911	54.165	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	108.340%		
Target Compounds					Qvalue	
13) Acetone	2.642	43	3002	2.961	ug/L	54


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090722\
Data File : VU050664.D
Acq On : 07 Sep 2022 16:38
Operator : SY/MD
Sample : N4482-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW346

Quant Time: Sep 08 05:43:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM090222WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 07 01:37:50 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 2.961 ug/L
 RT: 2.642 min Scan# 40
 Delta R.T. 0.020 min
 Lab File: VU050664.D
 Acq: 07 Sep 2022 16:38

Instrument :
 MSVOA_U
 ClientSampleId :
 EW346

Tgt Ion: 43 Resp: 3002
 Ion Ratio Lower Upper
 43 100
 58 10.3 0.0 76.2

