

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
 Data File : VU055180.D
 Acq On : 07 Sep 2023 11:42
 Operator : MD/SY
 Sample : VU0907WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK029

Quant Time: Sep 08 05:05:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 08 05:04:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	161820	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	170062	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	79085	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	45154	5.954	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	119.000%	
7) Chloroethane-d5	1.907	69	48447	5.518	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	110.400%	
11) 1,1-Dichloroethene-d2	2.563	65	22635	5.238	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	104.800%	
20) 2-Butanone-d5	4.621	46	126589	57.518	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	115.040%	
24) Chloroform-d	5.062	84	108282	5.380	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	107.600%	
26) 1,2-Dichloroethane-d4	5.701	65	58135	5.774	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.400%	
32) Benzene-d6	5.724	84	219009	5.217	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.400%	
36) 1,2-Dichloropropane-d6	6.689	67	71435	5.350	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.000%	
41) Toluene-d8	7.898	98	189174	4.892	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.800%	
43) trans-1,3-Dichloroprop...	8.180	79	16386	4.457	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.200%	
46) 2-Hexanone-d5	8.634	63	76113	50.989	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.980%	
56) 1,1,2,2-Tetrachloroeth...	10.756	84	51660	5.502	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	110.000%	
66) 1,2-Dichlorobenzene-d4	12.193	152	64210	5.562	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.200%	
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	6222	0.544	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
Data File : VU055180.D
Acq On : 07 Sep 2023 11:42
Operator : MD/SY
Sample : VU0907WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK029

Quant Time: Sep 08 05:05:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 08 05:04:57 2023
Response via : Initial Calibration

