

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
 Data File : VU055182.D
 Acq On : 07 Sep 2023 12:37
 Operator : MD/SY
 Sample : 04230-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 08 05:06:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 08 05:04:57 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	159865	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	168190	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	81927	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	40444	5.398	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 108.000%		
7) Chloroethane-d5	1.907	69	45868	5.289	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 105.800%		
11) 1,1-Dichloroethene-d2	2.563	65	21383	5.009	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 100.200%		
20) 2-Butanone-d5	4.624	46	119133	54.792	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 109.580%		
24) Chloroform-d	5.062	84	100946	5.077	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 101.600%		
26) 1,2-Dichloroethane-d4	5.701	65	51539	5.181	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 103.600%		
32) Benzene-d6	5.724	84	199247	4.799	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 96.000%		
36) 1,2-Dichloropropane-d6	6.689	67	68176	5.163	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 103.200%		
41) Toluene-d8	7.898	98	174457	4.562	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 91.200%		
43) trans-1,3-Dichloroprop...	8.180	79	16387	4.506	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 90.200%		
46) 2-Hexanone-d5	8.634	63	68785	46.593	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 93.180%		
56) 1,1,2,2-Tetrachloroeth...	10.756	84	47352	5.099	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 102.000%		
66) 1,2-Dichlorobenzene-d4	12.193	152	60623	5.070	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 101.400%		
Target Compounds						
16) Methylene chloride	3.042	84	3631	0.322	ug/L	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
Data File : VU055182.D
Acq On : 07 Sep 2023 12:37
Operator : MD/SY
Sample : 04230-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 08 05:06:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 08 05:04:57 2023
Response via : Initial Calibration

