

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
 Data File : VU055204.D
 Acq On : 07 Sep 2023 22:56
 Operator : MD/SY
 Sample : VU0907WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK030

Quant Time: Sep 08 06:19:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 08 06:18:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	155327	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	160853	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	76071	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	38497	5.288	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.800%
7) Chloroethane-d5	1.910	69	44600	5.293	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.800%
11) 1,1-Dichloroethene-d2	2.563	65	21817	5.260	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.200%
20) 2-Butanone-d5	4.627	46	132948	62.932	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.860%
24) Chloroform-d	5.061	84	101755	5.267	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
26) 1,2-Dichloroethane-d4	5.701	65	55414	5.734	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.600%
32) Benzene-d6	5.724	84	203999	5.138	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
36) 1,2-Dichloropropane-d6	6.688	67	68487	5.423	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.400%
41) Toluene-d8	7.897	98	171311	4.684	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
43) trans-1,3-Dichloroprop...	8.180	79	15941	4.584	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.600%
46) 2-Hexanone-d5	8.634	63	69424	49.171	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.340%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	48767	5.491	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.800%
66) 1,2-Dichlorobenzene-d4	12.193	152	58458	5.265	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.200%
Target Compounds						
16) Methylene chloride	3.042	84	3885	0.354	ug/L	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
Data File : VU055204.D
Acq On : 07 Sep 2023 22:56
Operator : MD/SY
Sample : VU0907WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK030

Quant Time: Sep 08 06:19:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 08 06:18:43 2023
Response via : Initial Calibration

