

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
 Data File : VU055212.D
 Acq On : 08 Sep 2023 02:08
 Operator : MD/SY
 Sample : 04307-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4H6

Quant Time: Sep 08 06:39:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 08 06:18:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	175565	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.419	117	180294	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	90416	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	26784	3.255	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.200%
7) Chloroethane-d5	1.911	69	37370	3.923	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.400%
11) 1,1-Dichloroethene-d2	2.563	65	17856	3.809	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.200%
20) 2-Butanone-d5	4.628	46	138554	58.025	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.060%
24) Chloroform-d	5.062	84	101901	4.667	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	5.698	65	53091	4.860	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.200%
32) Benzene-d6	5.724	84	195172	4.386	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
36) 1,2-Dichloropropane-d6	6.689	67	65972	4.661	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%
41) Toluene-d8	7.898	98	165778	4.044	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
43) trans-1,3-Dichloroprop...	8.181	79	14918	3.827	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.600%
46) 2-Hexanone-d5	8.634	63	71828	45.388	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.780%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	51452	5.169	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.400%
66) 1,2-Dichlorobenzene-d4	12.193	152	62181	4.712	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.200%
Target Compounds						Qvalue
3) Chloromethane	1.519	50	1641	0.158	ug/L	85
5) Vinyl chloride	1.599	62	2399	0.219	ug/L	100
17) Methyl tert-butyl Ether	3.364	73	6914	0.325	ug/L #	89
18) trans-1,2-Dichloroethene	3.354	96	5781	0.702	ug/L	82
19) 1,1-Dichloroethane	3.866	63	7950	0.447	ug/L #	91
22) cis-1,2-Dichloroethene	4.663	96	219678	21.792	ug/L	96
25) Chloroform	5.081	83	9526	0.514	ug/L	91
31) Carbon tetrachloride	5.531	117	2550	0.197	ug/L #	80
34) Trichloroethene	6.541	95	111785	10.482	ug/L	98
45) 1,1,2-Trichloroethane	8.399	97	1192	0.169	ug/L #	66
47) Tetrachloroethene	8.554	164	50515	6.854	ug/L	98
70) 1,2,4-trichlorobenzene	13.840	180	29820	2.079	ug/L	97
72) 1,2,3-Trichlorobenzene	14.332	180	11085	0.890	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
Data File : VU055212.D
Acq On : 08 Sep 2023 02:08
Operator : MD/SY
Sample : 04307-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4H6

Quant Time: Sep 08 06:39:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 08 06:18:43 2023
Response via : Initial Calibration

