

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
 Data File : VU055215.D
 Acq On : 08 Sep 2023 03:20
 Operator : MD/SY
 Sample : 04307-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4J8

Quant Time: Sep 08 06:40:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 08 06:18:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	169454	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	180921	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	89765	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	29428	3.705	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.200%
7) Chloroethane-d5	1.908	69	38842	4.225	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.600%
11) 1,1-Dichloroethene-d2	2.563	65	17865	3.948	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.000%
20) 2-Butanone-d5	4.628	46	141027	61.191	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.380%
24) Chloroform-d	5.059	84	103769	4.924	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
26) 1,2-Dichloroethane-d4	5.698	65	54532	5.172	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.400%
32) Benzene-d6	5.724	84	197617	4.425	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.689	67	68649	4.833	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.898	98	168953	4.107	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	8.181	79	14680	3.753	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.000%
46) 2-Hexanone-d5	8.634	63	67634	42.589	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.180%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	52810	5.287	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	12.193	152	61994	4.732	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.600%
Target Compounds						
					Qvalue	
3) Chloromethane	1.515	50	4236	0.423	ug/L #	75
5) Vinyl chloride	1.599	62	1252	0.118	ug/L #	44
22) cis-1,2-Dichloroethene	4.663	96	96588	9.927	ug/L	94
31) Carbon tetrachloride	5.522	117	27004	2.075	ug/L	95
34) Trichloroethene	6.541	95	51307	4.794	ug/L	95
47) Tetrachloroethene	8.554	164	17460	2.361	ug/L	97
70) 1,2,4-trichlorobenzene	13.846	180	13509	0.949	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090723\
Data File : VU055215.D
Acq On : 08 Sep 2023 03:20
Operator : MD/SY
Sample : 04307-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4J8

Quant Time: Sep 08 06:40:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 08 06:18:43 2023
Response via : Initial Calibration

