

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090818\
 Data File : VU026627.D
 Acq On : 07 Sep 2018 18:47
 Operator : MD/SY
 Sample : J4749-17
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E5FB8

Quant Time: Sep 08 02:08:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 08 01:21:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	141434	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	129842	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	50095	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	41919	47.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.68%
7) Chloroethane-d5	1.68	69	39256	49.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.27	63	62066	31.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.90%
21) 2-Butanone-d5	4.18	46	78246	115.05	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	115.05%
24) Chloroform-d	4.65	84	82048	49.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.14%
26) 1,2-Dichloroethane-d4	5.31	65	55870	50.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.40%
32) Benzene-d6	5.34	84	158522	52.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.94%
36) 1,2-Dichloropropane-d6	6.33	67	56886	55.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.66%
41) Toluene-d8	7.57	98	133492	46.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.42%
43) trans-1,3-Dichloropropene-	7.85	79	23232	49.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.98%
47) 2-Hexanone-d5	8.31	63	53583	113.64	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.64%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	79114	52.23	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.46%
64) 1,2-Dichlorobenzene-d4	11.86	152	50658	55.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.48%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090818\
Data File : VU026627.D
Acq On : 07 Sep 2018 18:47
Operator : MD/SY
Sample : J4749-17
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E5FB8

Quant Time: Sep 08 02:08:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082918WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 08 01:21:38 2018
Response via : Initial Calibration

