

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090819\
 Data File : VU034379.D
 Acq On : 07 Sep 2019 14:37
 Operator : JC/SP
 Sample : K4682-12
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0BN5

Quant Time: Sep 09 05:21:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR090119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 03 14:18:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	301504	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	300247	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	148197	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	78240	3.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.00%
7) Chloroethane-d5	1.68	69	71008	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.20%
11) 1,1-Dichloroethene-d2	2.26	63	116426	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	4.17	46	234805	50.05	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.10%
24) Chloroform-d	4.62	84	162478	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.29	65	82634	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	5.32	84	349266	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.40%
36) 1,2-Dichloropropane-d6	6.31	67	111951	4.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.00%
41) Toluene-d8	7.55	98	307013	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.60%
43) trans-1,3-Dichloropropene-	7.83	79	37728	4.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.20%
46) 2-Hexanone-d5	8.30	63	180276	48.11	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.22%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	90939	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	11.84	152	130525	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090819\
Data File : VU034379.D
Acq On : 07 Sep 2019 14:37
Operator : JC/SP
Sample : K4682-12
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0BN5

Quant Time: Sep 09 05:21:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR090119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 03 14:18:00 2019
Response via : Initial Calibration

