

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090820\
 Data File : VU040053.D
 Acq On : 08 Sep 2020 15:18
 Operator : SY/MD
 Sample : L3933-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Sep 09 03:09:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 09 01:36:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	119023	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	119638	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	54713	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	26726	4.11	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.20%
7) Chloroethane-d5	1.92	69	24516	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.60%
11) 1,1-Dichloroethene-d2	2.58	63	48447	2.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.20%#
20) 2-Butanone-d5	4.64	46	135781	52.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.48%
24) Chloroform-d	5.08	84	70942	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	5.71	65	44568	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
32) Benzene-d6	5.74	84	127669	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.70	67	43728	4.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.00%
41) Toluene-d8	7.90	98	107093	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.80%
43) trans-1,3-Dichloropropene-	8.19	79	18250	4.31	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.20%
46) 2-Hexanone-d5	8.64	63	97849	48.84	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.68%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	39372	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	45175	5.10	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090820\
Data File : VU040053.D
Acq On : 08 Sep 2020 15:18
Operator : SY/MD
Sample : L3933-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Sep 09 03:09:48 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 09 01:36:31 2020
Response via : Initial Calibration

