

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090821\
 Data File : VU044602.D
 Acq On : 08 Sep 2021 11:32
 Operator : SY/MD
 Sample : M3622-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 09 07:36:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 09 07:35:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	113705	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	117494	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	58548	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	33976	3.695	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	73.800%	
7) Chloroethane-d5	1.919	69	35359	4.424	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	88.400%	
11) 1,1-Dichloroethene-d2	2.572	65	15370	3.929	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	78.600%	
20) 2-Butanone-d5	4.645	46	74989	29.485	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	58.960%	
24) Chloroform-d	5.073	84	66915	4.115	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.400%	
26) 1,2-Dichloroethane-d4	5.710	65	36918	4.163	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.200%	
32) Benzene-d6	5.735	84	139206	4.207	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.200%	
36) 1,2-Dichloropropane-d6	6.697	67	41791	3.948	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	79.000%	
41) Toluene-d8	7.906	98	120789	4.117	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.400%	
43) trans-1,3-Dichloroprop...	8.189	79	14389	3.785	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.800%	
46) 2-Hexanone-d5	8.645	63	87049	42.348	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	84.700%	
56) 1,1,2,2-Tetrachloroeth...	10.764	84	38645	4.335	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	86.800%	
66) 1,2-Dichlorobenzene-d4	12.198	152	47713	4.823	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.400%	
Target Compounds						Qvalue
16) Methylene chloride	3.051	84	11681	0.900	ug/L	88
25) Chloroform	5.096	83	3680	0.203	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090821\
Data File : VU044602.D
Acq On : 08 Sep 2021 11:32
Operator : SY/MD
Sample : M3622-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 09 07:36:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 09 07:35:45 2021
Response via : Initial Calibration









#16
Methylene chloride
Concen: 0.900 ug/L
RT: 3.051 min Scan# 532
Delta R.T. -0.003 min
Lab File: VU044602.D
Acq: 08 Sep 2021 11:32

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Tgt Ion: 84 Resp: 11681

Ion	Ratio	Lower	Upper
84	100		
86	63.2	45.2	84.0
49	107.2	88.8	165.0



#25
Chloroform
Concen: 0.203 ug/L
RT: 5.096 min Scan# 1168
Delta R.T. 0.000 min
Lab File: VU044602.D
Acq: 08 Sep 2021 11:32

Tgt Ion: 83 Resp: 3680

Ion	Ratio	Lower	Upper
83	100		
85	67.0	45.3	84.1

