

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090919\
 Data File : VU034418.D
 Acq On : 09 Sep 2019 09:23
 Operator : JC/SP
 Sample : VU0909WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0909WBL01

Quant Time: Sep 10 06:04:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	256784	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	362842	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	370952	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	209971	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	208154	53.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.16%	
35) Dibromofluoromethane	4.86	113	184745	54.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.44%	
50) Toluene-d8	7.54	98	634781	50.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.08%	
62) 4-Bromofluorobenzene	10.29	95	209380	42.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.50%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090919\
Data File : VU034418.D
Acq On : 09 Sep 2019 09:23
Operator : JC/SP
Sample : VU0909WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0909WBL01

Quant Time: Sep 10 06:04:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

