

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090919\
 Data File : VU034420.D
 Acq On : 09 Sep 2019 10:07
 Operator : JC/SP
 Sample : PB122905TB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB122905TB

Quant Time: Sep 10 06:09:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	244156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	363509	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	404548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	212474	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	193990	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
35) Dibromofluoromethane	4.86	113	172412	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
50) Toluene-d8	7.55	98	620423	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
62) 4-Bromofluorobenzene	10.29	95	209181	42.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.26%	
Target Compounds						
16) Acetone	2.31	43	9795	5.124	ug/l	94
20) Methylene Chloride	2.69	84	6192	1.591	ug/l	97
43) Isopropyl Acetate	5.53	43	67055	9.222	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090919\
Data File : VU034420.D
Acq On : 09 Sep 2019 10:07
Operator : JC/SP
Sample : PB122905TB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB122905TB

Quant Time: Sep 10 06:09:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

