

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090919\
 Data File : VU034421.D
 Acq On : 09 Sep 2019 10:30
 Operator : JC/SP
 Sample : K4601-38
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T12-F4-(2.7)

Quant Time: Sep 10 06:12:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	251030	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	375986	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	405401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	214185	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	210472	55.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.82%	
35) Dibromofluoromethane	4.86	113	186541	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
50) Toluene-d8	7.55	98	678913	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
62) 4-Bromofluorobenzene	10.29	95	220692	43.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.98%	
Target Compounds						
16) Acetone	2.31	43	15779	8.028	ug/l	100
20) Methylene Chloride	2.69	84	13827	3.927	ug/l	97
43) Isopropyl Acetate	5.53	43	46264	6.151	ug/l	99
76) 1,2,3-Trichloropropane	10.29	75	102157	21.850	ug/l #	36

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090919\
Data File : VU034421.D
Acq On : 09 Sep 2019 10:30
Operator : JC/SP
Sample : K4601-38
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T12-F4-(2.7)

Quant Time: Sep 10 06:12:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

