

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090919\
 Data File : VU034428.D
 Acq On : 09 Sep 2019 13:13
 Operator : JC/SP
 Sample : K4602-22
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T12-G3-(0)

Quant Time: Sep 10 06:22:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	261461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	387168	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	415416	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	222641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	211310	53.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.84%	
35) Dibromofluoromethane	4.86	113	186314	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
50) Toluene-d8	7.54	98	678343	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
62) 4-Bromofluorobenzene	10.29	95	227701	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
16) Acetone	2.31	43	13926	6.802	ug/l	99
20) Methylene Chloride	2.69	84	5911	1.374	ug/l	96
43) Isopropyl Acetate	5.53	43	52511	6.780	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090919\
Data File : VU034428.D
Acq On : 09 Sep 2019 13:13
Operator : JC/SP
Sample : K4602-22
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T12-G3-(0)

Quant Time: Sep 10 06:22:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

