

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU090919\
 Data File : VU034432.D
 Acq On : 09 Sep 2019 14:46
 Operator : JC/SP
 Sample : K4602-25
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T12-G2-(2.7)

Quant Time: Sep 10 06:49:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	255878	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	386177	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	400059	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	222674	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	212241	54.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.64%	
35) Dibromofluoromethane	4.86	113	186431	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
50) Toluene-d8	7.54	98	675137	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	10.29	95	232220	44.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.10%	
Target Compounds						
16) Acetone	2.31	43	14657	7.316	ug/l	89
20) Methylene Chloride	2.69	84	8577	2.232	ug/l	98
43) Isopropyl Acetate	5.53	43	55406	7.173	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU090919\
Data File : VU034432.D
Acq On : 09 Sep 2019 14:46
Operator : JC/SP
Sample : K4602-25
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T12-G2-(2.7)

Quant Time: Sep 10 06:49:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

