

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090919\
 Data File : VU034433.D
 Acq On : 09 Sep 2019 15:09
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 10 06:50:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.523	0.481	8.0	84	0.00
3 P	Chloromethane	0.644	0.534	17.1	80	0.00
4 C	Vinyl Chloride	0.652	0.604	7.4#	86	0.00
5 T	Bromomethane	0.360	0.361	-0.3	91	0.00
6 T	Chloroethane	0.427	0.387	9.4	85	0.00
7 T	Trichlorofluoromethane	0.874	0.861	1.5	90	0.00
8 T	Diethyl Ether	0.384	0.385	-0.3	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.579	0.566	2.2	90	0.00
10 T	Methyl Iodide	0.477	0.458	4.0	76	0.00
11 T	Tert butyl alcohol	0.160	0.168	-5.0	91	0.00
12 CM	1,1-Dichloroethene	0.544	0.531	2.4#	88	0.00
13 T	Acrolein	0.134	0.107	20.1#	70	0.00
14 T	Allyl chloride	1.040	0.990	4.8	86	0.00
15 T	Acrylonitrile	0.398	0.414	-4.0	87	0.00
16 T	Acetone	0.391	0.359	8.2	84	0.00
17 T	Carbon Disulfide	1.223	1.036	15.3	79	0.00
18 T	Methyl Acetate	0.953	0.914	4.1	88	0.00
19 T	Methyl tert-butyl Ether	1.983	2.074	-4.6	90	0.00
20 T	Methylene Chloride	0.709	0.686	3.2	90	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.563	1.7	89	0.00
22 T	Diisopropyl ether	2.028	2.110	-4.0	88	0.00
23 T	Vinyl Acetate	1.652	1.735	-5.0	87	0.00
24 P	1,1-Dichloroethane	1.211	1.195	1.3	89	0.00
25 T	2-Butanone	0.523	0.545	-4.2	86	0.00
26 T	2,2-Dichloropropane	0.974	0.966	0.8	89	0.00
27 T	cis-1,2-Dichloroethene	0.696	0.718	-3.2	92	0.00
28 T	Bromochloromethane	0.557	0.558	-0.2	82	0.00
29 T	Tetrahydrofuran	0.318	0.335	-5.3	86	0.00
30 C	Chloroform	1.231	1.216	1.2#	90	0.00
31 T	Cyclohexane	0.886	0.834	5.9	85	0.00
32 T	1,1,1-Trichloroethane	0.984	1.002	-1.8	92	0.00
33 S	1,2-Dichloroethane-d4	0.757	0.743	1.8	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.465	0.470	-1.1	87	0.00
36 T	1,1-Dichloropropene	0.542	0.536	1.1	89	0.00
37 T	Ethyl Acetate	0.641	0.656	-2.3	87	0.00
38 T	Carbon Tetrachloride	0.579	0.576	0.5	89	0.00
39 T	Methylcyclohexane	0.580	0.575	0.9	86	0.00
40 TM	Benzene	1.750	1.769	-1.1	90	0.00
41 T	Methacrylonitrile	0.342	0.366	-7.0	87	0.00
42 TM	1,2-Dichloroethane	0.607	0.604	0.5	89	0.00
43 T	Isopropyl Acetate	1.000	1.028	-2.8	88	0.00
44 TM	Trichloroethene	0.484	0.489	-1.0	91	0.00
45 C	1,2-Dichloropropane	0.512	0.517	-1.0#	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090919\
 Data File : VU034433.D
 Acq On : 09 Sep 2019 15:09
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 10 06:50:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.321	0.332	-3.4	91	0.00
47 T	Bromodichloromethane	0.661	0.676	-2.3	91	0.00
48 T	Methyl methacrylate	0.475	0.480	-1.1	85	0.00
49 T	1,4-Dioxane	0.010	0.012	-20.0	94	0.00
50 S	Toluene-d8	1.748	1.781	-1.9	83	0.00
51 T	4-Methyl-2-Pentanone	0.692	0.714	-3.2	87	0.00
52 CM	Toluene	1.098	1.135	-3.4#	89	0.00
53 T	t-1,3-Dichloropropene	0.685	0.717	-4.7	91	0.00
54 T	cis-1,3-Dichloropropene	0.752	0.793	-5.5	90	0.00
55 T	1,1,2-Trichloroethane	0.498	0.512	-2.8	91	0.00
56 T	Ethyl methacrylate	0.672	0.727	-8.2	88	0.00
57 T	1,3-Dichloropropane	0.812	0.827	-1.8	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.222	0.260	-17.1	86	0.00
59 T	2-Hexanone	0.517	0.542	-4.8	86	0.00
60 T	Dibromochloromethane	0.535	0.566	-5.8	92	0.00
61 T	1,2-Dibromoethane	0.488	0.513	-5.1	91	0.00
62 S	4-Bromofluorobenzene	0.675	0.681	-0.9	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.434	0.432	0.5	88	0.00
65 PM	Chlorobenzene	1.269	1.337	-5.4	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.484	0.508	-5.0	92	0.00
67 C	Ethyl Benzene	2.074	2.226	-7.3#	89	0.00
68 T	m/p-Xylenes	0.781	0.879	-12.5	90	0.00
69 T	o-Xylene	0.764	0.850	-11.3	90	0.00
70 T	Styrene	1.339	1.558	-16.4	92	0.00
71 P	Bromoform	0.420	0.462	-10.0	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.254	3.283	-0.9	90	0.00
74 T	N-amyl acetate	1.310	1.374	-4.9	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.317	1.267	3.8	91	0.00
76 T	1,2,3-Trichloropropane	1.091	1.041	4.6	91	0.00
77 T	Bromobenzene	0.910	0.911	-0.1	93	0.00
78 T	n-propylbenzene	3.758	3.864	-2.8	91	0.00
79 T	2-Chlorotoluene	2.293	2.298	-0.2	91	0.00
80 T	1,3,5-Trimethylbenzene	2.721	2.933	-7.8	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.325	0.345	-6.2	93	0.00
82 T	4-Chlorotoluene	2.639	2.684	-1.7	91	0.00
83 T	tert-Butylbenzene	2.722	2.837	-4.2	93	0.00
84 T	1,2,4-Trimethylbenzene	2.739	2.975	-8.6	93	0.00
85 T	sec-Butylbenzene	3.214	3.403	-5.9	92	0.00
86 T	p-Isopropyltoluene	2.966	3.224	-8.7	94	0.00
87 T	1,3-Dichlorobenzene	1.660	1.712	-3.1	96	0.00
88 T	1,4-Dichlorobenzene	1.732	1.743	-0.6	96	0.00
89 T	n-Butylbenzene	2.578	2.829	-9.7	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090919\
Data File : VU034433.D
Acq On : 09 Sep 2019 15:09
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Sep 10 06:50:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.475	0.486	-2.3	95	0.00
91 T	1,2-Dichlorobenzene	1.632	1.729	-5.9	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.242	0.269	-11.2	95	0.00
93 T	1,2,4-Trichlorobenzene	0.964	1.226	-27.2#	99	0.00
94 T	Hexachlorobutadiene	0.571	0.568	0.5	92	0.00
95 T	Naphthalene	2.598	3.342	-28.6#	95	0.00
96 T	1,2,3-Trichlorobenzene	1.004	1.222	-21.7#	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6