


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU090920\
 Data File : VU040057.D
 Acq On : 09 Sep 2020 09:58
 Operator : SY/MD
 Sample : BFB
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U090920DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Fri Sep 11 12:52:59 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.7	40930	PASS
75	95	30	60	57.0	102717	PASS
95	95	100	100	100.0	180352	PASS
96	95	5	9	6.6	11838	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	66.2	119392	PASS
175	174	5	9	7.6	9082	PASS
176	174	95	101	96.3	114928	PASS
177	176	5	9	6.3	7197	PASS