

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090924\
 Data File : VU060651.D
 Acq On : 09 Sep 2024 11:45
 Operator : MD/SY
 Sample : P3867-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Sep 09 23:45:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 05:03:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	4862	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	7314	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2098	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	3531	0.446	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	90.000%
4) 1,2-Dichloroethane-d4	5.685	65	2153	0.514	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
7) 1,2-Dichloropropane-d6	6.687	67	2318	0.510	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.887	98	5205	0.554	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	110.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1828	0.459	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	92.000%
Target Compounds						
13) 1,2-Dibromo-3-chloropr...	12.750	75	33	0.070	ug/L	Qvalue 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090924\
Data File : VU060651.D
Acq On : 09 Sep 2024 11:45
Operator : MD/SY
Sample : P3867-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :


Quant Time: Sep 09 23:45:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 05:03:57 2024
Response via : Initial Calibration





#13
 1,2-Dibromo-3-chloropropane
 Concen: 0.070 ug/L
 RT: 12.750 min Scan# 81
 Delta R.T. -0.240 min
 Lab File: VU060651.D
 Acq: 09 Sep 2024 11:45

Instrument :
 MSVOA_U
 ClientSampleId :



Tgt Ion: 75	Resp: 33
Ion Ratio	Lower Upper
75 100	
157 78.3	74.3 111.5
155 81.7	66.2 99.2

