

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090924\
 Data File : VU060658.D
 Acq On : 09 Sep 2024 14:49
 Operator : MD/SY
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5096

Quant Time: Sep 09 23:46:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 05:03:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	4637	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	6966	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2020	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	3420	0.453	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	90.000%		
4) 1,2-Dichloroethane-d4	5.685	65	2029	0.508	ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery =	102.000%		
7) 1,2-Dichloropropane-d6	6.687	67	2122	0.490	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	98.000%		
8) Toluene-d8	7.887	98	4846	0.542	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	108.000%		
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1696	0.447	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	90.000%		
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.597	62	3772	0.476	ug/L	100
6) Trichloroethene	6.544	95	1980	0.450	ug/L	96
9) 1,2-Dibromoethane	8.924	107	1317	0.477	ug/L	99
12) 1,2,3-Trichloropropane	10.814	75	1208	0.515	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.990	75	226	0.497	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU090924\
Data File : VU060658.D
Acq On : 09 Sep 2024 14:49
Operator : MD/SY
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5096

Quant Time: Sep 09 23:46:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 05:03:57 2024
Response via : Initial Calibration

