

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091019\
 Data File : VU034446.D
 Acq On : 10 Sep 2019 13:15
 Operator : JC/SP
 Sample : K4602-31
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PAR-C3-(3.11)

Quant Time: Sep 11 06:41:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	239462	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	377045	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	390371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	211148	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	204481	56.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.88%	
35) Dibromofluoromethane	4.86	113	179417	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
50) Toluene-d8	7.54	98	652873	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	10.29	95	222788	43.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.56%	
Target Compounds						
16) Acetone	2.31	43	11962	6.380	ug/l	96
20) Methylene Chloride	2.69	84	7166	1.949	ug/l	92
43) Isopropyl Acetate	5.53	43	56193	7.451	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091019\
Data File : VU034446.D
Acq On : 10 Sep 2019 13:15
Operator : JC/SP
Sample : K4602-31
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PAR-C3-(3.11)

Quant Time: Sep 11 06:41:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

