

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091019\
 Data File : VU034449.D
 Acq On : 10 Sep 2019 14:25
 Operator : JC/SP
 Sample : K4602-37
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PAR-D4-(0)

Quant Time: Sep 11 06:46:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	220390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	346943	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	359992	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	187770	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	195622	58.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.32%	
35) Dibromofluoromethane	4.86	113	168765	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
50) Toluene-d8	7.55	98	584503	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
62) 4-Bromofluorobenzene	10.29	95	194736	41.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.16%	
Target Compounds						
16) Acetone	2.31	43	11532	6.683	ug/l	97
20) Methylene Chloride	2.69	84	6368	1.868	ug/l	94
43) Isopropyl Acetate	5.53	43	56801	8.185	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091019\
Data File : VU034449.D
Acq On : 10 Sep 2019 14:25
Operator : JC/SP
Sample : K4602-37
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PAR-D4-(0)

Quant Time: Sep 11 06:46:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

