

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091019\
 Data File : VU034456.D
 Acq On : 10 Sep 2019 17:07
 Operator : JC/SP
 Sample : K4603-20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PAR-D4-(3.11)

Quant Time: Sep 11 06:57:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	223103	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	355934	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	365798	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	187160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	194061	57.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.98%	
35) Dibromofluoromethane	4.86	113	167546	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
50) Toluene-d8	7.55	98	566969	45.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.12%	
62) 4-Bromofluorobenzene	10.29	95	189790	39.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.00%	
Target Compounds						
16) Acetone	2.31	43	11437	6.547	ug/l	96
20) Methylene Chloride	2.69	84	11829	3.765	ug/l	98
43) Isopropyl Acetate	5.53	43	57626	8.094	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091019\
Data File : VU034456.D
Acq On : 10 Sep 2019 17:07
Operator : JC/SP
Sample : K4603-20
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PAR-D4-(3.11)

Quant Time: Sep 11 06:57:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

