

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091019\
 Data File : VU034457.D
 Acq On : 10 Sep 2019 17:30
 Operator : JC/SP
 Sample : K4602-36
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PAR-D3-(0)

Quant Time: Sep 11 07:01:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	183331	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	305288	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	314590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	152615	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	195214	70.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	140.76%	
35) Dibromofluoromethane	4.86	113	165711	58.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.66%	
50) Toluene-d8	7.55	98	502254	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
62) 4-Bromofluorobenzene	10.29	95	166267	40.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.70%	
Target Compounds						
16) Acetone	2.31	43	11495	8.008	ug/l	98
20) Methylene Chloride	2.69	84	7244	2.703	ug/l	93
43) Isopropyl Acetate	5.53	43	58563	9.590	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091019\
Data File : VU034457.D
Acq On : 10 Sep 2019 17:30
Operator : JC/SP
Sample : K4602-36
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PAR-D3-(0)

Quant Time: Sep 11 07:01:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

