

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091019\
 Data File : VU034461.D
 Acq On : 10 Sep 2019 19:03
 Operator : JC/SP
 Sample : K4603-25
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PAR-E4-(0)

Quant Time: Sep 11 07:08:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	235247	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	367734	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	382119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	208519	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	202076	56.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.54%	
35) Dibromofluoromethane	4.86	113	177433	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
50) Toluene-d8	7.55	98	643297	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	10.29	95	217214	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
Target Compounds						
16) Acetone	2.32	43	10945	5.942	ug/l	100
20) Methylene Chloride	2.69	84	14405	4.410	ug/l	96
43) Isopropyl Acetate	5.53	43	63873	8.683	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091019\
Data File : VU034461.D
Acq On : 10 Sep 2019 19:03
Operator : JC/SP
Sample : K4603-25
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PAR-E4-(0)

Quant Time: Sep 11 07:08:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

