

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091019\
 Data File : VU034462.D
 Acq On : 10 Sep 2019 19:26
 Operator : JC/SP
 Sample : K4601-35
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T12-F1-(2.7)

Quant Time: Sep 11 07:10:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	225973	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	367685	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	393068	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	213795	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	215208	62.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.88%	
35) Dibromofluoromethane	4.86	113	185254	54.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.30%	
50) Toluene-d8	7.55	98	656442	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	10.29	95	223273	44.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.98%	
Target Compounds						
16) Acetone	2.32	43	14802	8.366	ug/l	95
20) Methylene Chloride	2.69	84	7381	2.165	ug/l	94
43) Isopropyl Acetate	5.53	43	51243	6.967	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091019\
Data File : VU034462.D
Acq On : 10 Sep 2019 19:26
Operator : JC/SP
Sample : K4601-35
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T12-F1-(2.7)

Quant Time: Sep 11 07:10:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

