

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
 Data File : VU060671.D
 Acq On : 10 Sep 2024 17:04
 Operator : MD/SY
 Sample : VIBLK132
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK132

Quant Time: Sep 11 02:49:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 11 02:46:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	95062	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	103956	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	44508	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	22437	5.344	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.800%
7) Chloroethane-d5	1.911	69	23694	3.725	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.400%
11) 1,1-Dichloroethene-d2	2.560	65	11671	4.716	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.400%
20) 2-Butanone-d5	4.634	46	69646	48.237	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.480%
24) Chloroform-d	5.055	84	64786	5.412	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.200%
26) 1,2-Dichloroethane-d4	5.695	65	34158	5.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.400%
32) Benzene-d6	5.721	84	114711	5.046	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
36) 1,2-Dichloropropane-d6	6.685	67	39312	5.360	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.200%
41) Toluene-d8	7.891	98	96687	4.534	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
43) trans-1,3-Dichloroprop...	8.177	79	11673	4.448	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	8.631	63	47604	37.883	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.760%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	29897	4.581	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	35764	5.450	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.000%
Target Compounds						
16) Methylene chloride	3.039	84	5197	0.687	ug/L	Qvalue 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
Data File : VU060671.D
Acq On : 10 Sep 2024 17:04
Operator : MD/SY
Sample : VIBLK132
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK132

Quant Time: Sep 11 02:49:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 11 02:46:13 2024
Response via : Initial Calibration

