

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
 Data File : VU060665.D
 Acq On : 10 Sep 2024 14:39
 Operator : MD/SY
 Sample : P3918-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAN6

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2024
 Supervised By :Mahesh Dadoda 09/11/2024

Quant Time: Sep 11 02:47:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 11 02:46:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	114190	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	122369	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	52577	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	25421	5.040	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	100.800%	
7) Chloroethane-d5	1.911	69	27929	3.655	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	73.000%	
11) 1,1-Dichloroethene-d2	2.560	65	12834	4.318	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	86.400%	
20) 2-Butanone-d5	4.634	46	73777	42.538	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	85.080%	
24) Chloroform-d	5.055	84	69762	4.852	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.000%	
26) 1,2-Dichloroethane-d4	5.698	65	37901	5.051	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.000%	
32) Benzene-d6	5.721	84	129630	4.844	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6	6.685	67	43653	5.056	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.200%	
41) Toluene-d8	7.891	98	109764	4.372	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.400%	
43) trans-1,3-Dichloroprop...	8.177	79	14377	4.654	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.000%	
46) 2-Hexanone-d5	8.631	63	55757	37.695	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	75.380%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	32738	4.261	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.200%	
66) 1,2-Dichlorobenzene-d4	12.187	152	39970	5.156	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.200%	
Target Compounds						Qvalue
10) 1,1,2-Trichloro-1,2,2-...	2.573	101	37899	3.727	ug/L	99
12) 1,1-Dichloroethene	2.576	96	4456	0.476	ug/L #	1
19) 1,1-Dichloroethane	3.866	63	2420	0.147	ug/L	88
22) cis-1,2-Dichloroethene	4.666	96	11965m	1.318	ug/L	
34) Trichloroethene	6.538	95	32499	3.089	ug/L	97
47) Tetrachloroethene	8.547	164	54141	6.279	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
Data File : VU060665.D
Acq On : 10 Sep 2024 14:39
Operator : MD/SY
Sample : P3918-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_U
ClientSampleId :
C0AN6

Quant Time: Sep 11 02:47:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 11 02:46:13 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2024
Supervised By :Mahesh Dadoda 09/11/2024









#10
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 3.727 ug/L
RT: 2.573 min Scan# 399
Delta R.T. -0.003 min
Lab File: VU060665.D
Acq: 10 Sep 2024 14:39

Instrument :
MSVOA_U
ClientSampleId :
COAN6

Tgt Ion:101 Resp: 3789
Ion Ratio Lower Upper
101 100
85 44.1 35.8 53.6
151 76.7 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 09/11/2024
Supervised By :Mahesh Dadoda 09/11/2024



#12
1,1-Dichloroethene
Concen: 0.476 ug/L
RT: 2.576 min Scan# 400
Delta R.T. -0.000 min
Lab File: VU060665.D
Acq: 10 Sep 2024 14:39

Tgt Ion: 96 Resp: 4456
Ion Ratio Lower Upper
96 100
61 165.1 116.4 216.2
63 382.2 74.8 139.0#





#19

1,1-Dichloroethane

Concen: 0.147 ug/L

RT: 3.866 min Scan# 800

Delta R.T. 0.003 min

Lab File: VU060665.D

Acq: 10 Sep 2024 14:39

Instrument :

MSVOA_U

ClientSampleId :





COAN6

Tgt Ion: 63	Resp: 2420
Ion Ratio	Lower Upper
63 100	
65 20.7	20.5 38.1
83 11.4	9.2 17.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2024
Supervised By :Mahesh Dadoda 09/11/2024



#22

cis-1,2-Dichloroethene

Concen: 1.318 ug/L m




RT: 4.666 min Scan# 1050


Delta R.T. 0.006 min

Lab File: VU060665.D

Acq: 10 Sep 2024 14:39


Tgt Ion: 96	Resp: 11965
Ion Ratio	Lower Upper
96 100	
61 128.6	92.9 172.5
98 61.6	45.0 83.6





#34
Trichloroethene
Concen: 3.089 ug/L
RT: 6.538 min Scan# 1632
Delta R.T. 0.003 min
Lab File: VU060665.D
Acq: 10 Sep 2024 14:39




Instrument :
MSVOA_U
ClientSampleId :
COAN6




Tgt Ion: 95 Resp: 32499
Ion Ratio Lower Upper
95 100
97 64.5 45.8 85.2
132 93.5 67.4 125.2
130 97.7 72.2 134.2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/11/2024
Supervised By :Mahesh Dadoda 09/11/2024



#47
Tetrachloroethene
Concen: 6.279 ug/L
RT: 8.547 min Scan# 2257
Delta R.T. -0.000 min
Lab File: VU060665.D
Acq: 10 Sep 2024 14:39



Tgt Ion:164 Resp: 54141
Ion Ratio Lower Upper
164 100
129 96.9 67.8 125.8
131 94.9 67.2 124.8
166 131.4 88.5 164.4

