

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
 Data File : VU060670.D
 Acq On : 10 Sep 2024 16:39
 Operator : MD/SY
 Sample : P3918-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AR3

Quant Time: Sep 11 02:49:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 11 02:46:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	6.242	114	99730	5.000 ug/L	0.00
28) Chlorobenzene-d5	9.412	117	105486	5.000 ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	44604	5.000 ug/L	0.00

System Monitoring Compounds					
4) Vinyl Chloride-d3	1.596	65	23848	5.414 ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 108.200%	
7) Chloroethane-d5	1.911	69	25183	3.773 ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 75.400%	
11) 1,1-Dichloroethene-d2	2.563	65	11927	4.594 ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 91.800%	
20) 2-Butanone-d5	4.634	46	69207	45.689 ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 91.380%	
24) Chloroform-d	5.055	84	65141	5.187 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.800%	
26) 1,2-Dichloroethane-d4	5.695	65	35534	5.422 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 108.400%	
32) Benzene-d6	5.721	84	120835	5.238 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 104.800%	
36) 1,2-Dichloropropane-d6	6.682	67	39009	5.242 ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 104.800%	
41) Toluene-d8	7.891	98	102936	4.757 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 95.200%	
43) trans-1,3-Dichloroprop...	8.177	79	12907	4.846 ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 97.000%	
46) 2-Hexanone-d5	8.631	63	49564	38.871 ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 77.740%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	31199	4.711 ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 94.200%	
66) 1,2-Dichlorobenzene-d4	12.187	152	37088	5.639 ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 112.800%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
Data File : VU060670.D
Acq On : 10 Sep 2024 16:39
Operator : MD/SY
Sample : P3918-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AR3

Quant Time: Sep 11 02:49:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 11 02:46:13 2024
Response via : Initial Calibration

