

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
 Data File : VU060672.D
 Acq On : 10 Sep 2024 17:28
 Operator : MD/SY
 Sample : P3918-07DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP5DL

Quant Time: Sep 11 02:49:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 11 02:46:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	97868	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	105811	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	44889	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	24753	5.726	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.600%
7) Chloroethane-d5	1.911	69	26921	4.111	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.200%
11) 1,1-Dichloroethene-d2	2.560	65	11858	4.655	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.000%
20) 2-Butanone-d5	4.631	46	73545	49.477	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.960%
24) Chloroform-d	5.055	84	71301	5.786	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.800%
26) 1,2-Dichloroethane-d4	5.695	65	37099	5.769	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.400%
32) Benzene-d6	5.721	84	128319	5.546	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.000%
36) 1,2-Dichloropropane-d6	6.685	67	42949	5.753	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.000%
41) Toluene-d8	7.891	98	106903	4.925	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
43) trans-1,3-Dichloroprop...	8.174	79	13589	5.087	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.800%
46) 2-Hexanone-d5	8.631	63	52881	41.345	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.680%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	32727	4.927	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	38647	5.839	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.800%
Target Compounds						
					Qvalue	
16) Methylene chloride	3.039	84	7468	0.959	ug/L	84
17) Methyl tert-butyl Ether	3.357	73	131714	8.091	ug/L	98
22) cis-1,2-Dichloroethene	4.660	96	16182	2.080	ug/L	98
34) Trichloroethene	6.537	95	4155	0.457	ug/L	92
47) Tetrachloroethene	8.547	164	21528	2.887	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091024\
Data File : VU060672.D
Acq On : 10 Sep 2024 17:28
Operator : MD/SY
Sample : P3918-07DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP5DL

Quant Time: Sep 11 02:49:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 11 02:46:13 2024
Response via : Initial Calibration









#16
Methylene chloride
Concen: 0.959 ug/L
RT: 3.039 min Scan# 544
Delta R.T. -0.003 min
Lab File: VU060672.D
Acq: 10 Sep 2024 17:28




Instrument :
MSVOA_U
ClientSampleId :
C0AP5DL


Tgt Ion: 84 Resp: 7468
Ion Ratio Lower Upper
84 100
86 61.8 47.5 88.3
49 111.4 94.4 175.2



#17
Methyl tert-butyl Ether
Concen: 8.091 ug/L
RT: 3.357 min Scan# 643
Delta R.T. -0.000 min
Lab File: VU060672.D
Acq: 10 Sep 2024 17:28

Tgt Ion: 73 Resp: 131714
Ion Ratio Lower Upper
73 100
43 19.3 15.8 23.6
57 22.1 16.8 25.2





#22

cis-1,2-Dichloroethene

Concen: 2.080 ug/L

RT: 4.660 min Scan# 1048

Delta R.T. -0.000 min

Lab File: VU060672.D

Acq: 10 Sep 2024 17:28

Instrument :

MSVOA_U

ClientSampleId :

C0AP5DL





Tgt Ion: 96 Resp: 16182

Ion Ratio Lower Upper

96 100

61 130.2 92.9 172.5

98 63.8 45.0 83.6



#34

Trichloroethene

Concen: 0.457 ug/L

RT: 6.537 min Scan# 1632

Delta R.T. 0.003 min

Lab File: VU060672.D

Acq: 10 Sep 2024 17:28

Tgt Ion: 95 Resp: 4155




Ion Ratio Lower Upper


95 100

97 57.4 45.8 85.2

132 83.3 67.4 125.2


130 101.4 72.2 134.2





#47
 Tetrachloroethene
 Concen: 2.887 ug/L
 RT: 8.547 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VU060672.D
 Acq: 10 Sep 2024 17:28

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP5DL



Tgt Ion:	164	Resp:	21528
Ion	Ratio	Lower	Upper
164	100		
129	90.4	67.8	125.8
131	88.6	67.2	124.8
166	116.3	88.5	164.4

