

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091118\
 Data File : VU026672.D
 Acq On : 11 Sep 2018 08:24
 Operator : MD/SY
 Sample : VU0911WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK51

Quant Time: Sep 12 00:48:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 12 00:46:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	190389	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	175159	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	73029	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	57818	40.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.50%
7) Chloroethane-d5	1.67	69	50603	44.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.52%
11) 1,1-Dichloroethene-d2	2.27	63	81496	32.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.20%
21) 2-Butanone-d5	4.18	46	84537	95.75	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.75%
24) Chloroform-d	4.65	84	100638	42.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.32%
26) 1,2-Dichloroethane-d4	5.31	65	71011	45.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.98%
32) Benzene-d6	5.34	84	215611	47.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.12%
36) 1,2-Dichloropropane-d6	6.33	67	74276	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.64%
41) Toluene-d8	7.57	98	190181	44.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.42%
43) trans-1,3-Dichloropropene-	7.85	79	32028	45.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.56%
47) 2-Hexanone-d5	8.31	63	59031	95.10	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.10%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	96741	46.44	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.88%
64) 1,2-Dichlorobenzene-d4	11.86	152	71497	49.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.22%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091118\
Data File : VU026672.D
Acq On : 11 Sep 2018 08:24
Operator : MD/SY
Sample : VU0911WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK51

Quant Time: Sep 12 00:48:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091018WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 12 00:46:21 2018
Response via : Initial Calibration

