

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091123\
 Data File : VU055238.D
 Acq On : 11 Sep 2023 16:28
 Operator : MD/SY
 Sample : 04354-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-01-090723

Quant Time: Sep 12 09:01:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U091123DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Sep 12 08:56:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.116	96	49721	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.637	95	15900	0.824	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
68) 1,2-Dichlorobenzene-d4	12.196	152	16059	0.854	ug/l	0.00
Spiked Amount 1.000			Recovery	=	85.000%	
Target Compounds						
27) Chloroform	5.087	83	54056	1.856	ug/l	95
44) Bromodichloromethane	7.110	83	6676	0.338	ug/l #	96
55) Dibromochloromethane	8.807	129	14797	1.220	ug/l	97
67) Bromoform	10.290	173	16579	2.703	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091123\
Data File : VU055238.D
Acq On : 11 Sep 2023 16:28
Operator : MD/SY
Sample : 04354-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-01-090723

Quant Time: Sep 12 09:01:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U091123DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Sep 12 08:56:31 2023
Response via : Initial Calibration

