

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091124\
 Data File : VU060694.D
 Acq On : 11 Sep 2024 19:20
 Operator : MD/SY
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5098

Quant Time: Sep 12 02:24:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 05:03:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	3802	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	5750	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1700	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	2765	0.447	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	90.000%		
4) 1,2-Dichloroethane-d4	5.685	65	1921	0.586	ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery =	118.000%		
7) 1,2-Dichloropropane-d6	6.687	67	1778	0.497	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	100.000%		
8) Toluene-d8	7.887	98	4210	0.570	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	114.000%		
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1442	0.461	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	92.000%		
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	2992	0.461	ug/L	99
6) Trichloroethene	6.544	95	1642	0.452	ug/L	96
9) 1,2-Dibromoethane	8.924	107	1073	0.471	ug/L #	98
12) 1,2,3-Trichloropropane	10.814	75	1054	0.534	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.990	75	204	0.533	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091124\
Data File : VU060694.D
Acq On : 11 Sep 2024 19:20
Operator : MD/SY
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5098

Quant Time: Sep 12 02:24:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 05:03:57 2024
Response via : Initial Calibration

