

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091124\
 Data File : VU060705.D
 Acq On : 12 Sep 2024 00:12
 Operator : MD/SY
 Sample : VIBLK010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK010

Quant Time: Sep 12 02:53:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 05:03:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	3641	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	5607	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	1604	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	2414	0.408	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	82.000%		
4) 1,2-Dichloroethane-d4	5.685	65	1705	0.543	ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery =	108.000%		
7) 1,2-Dichloropropane-d6	6.687	67	1595	0.458	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	92.000%		
8) Toluene-d8	7.887	98	3742	0.520	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	104.000%		
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1297	0.425	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	84.000%		
Target Compounds					Qvalue	
6) Trichloroethene	6.544	95	72	0.020	ug/L #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091124\
Data File : VU060705.D
Acq On : 12 Sep 2024 00:12
Operator : MD/SY
Sample : VIBLK010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK010

Quant Time: Sep 12 02:53:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 05:03:57 2024
Response via : Initial Calibration

