

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091218\
 Data File : VU026686.D
 Acq On : 11 Sep 2018 18:11
 Operator : MD/SY
 Sample : J4803-10ME
 Misc : 5.61g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP1-Q2-H5ME

Quant Time: Sep 13 09:27:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U091218W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 05:32:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	101080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	175968	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	182650	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	121345	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	84309	54.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.96%	
35) Dibromofluoromethane	4.88	113	73097	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
50) Toluene-d8	7.57	98	262912	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
62) 4-Bromofluorobenzene	10.31	95	122104	63.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.92%	
Target Compounds						
80) 1,3,5-Trimethylbenzene	10.78	105	177456	25.53	ug/l	Qvalue 100
86) p-Isopropyltoluene	11.48	119	65987	9.07	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091218\
Data File : VU026686.D
Acq On : 11 Sep 2018 18:11
Operator : MD/SY
Sample : J4803-10ME
Misc : 5.61µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP1-Q2-H5ME

Quant Time: Sep 13 09:27:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U091218W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 05:32:45 2018
Response via : Initial Calibration

