

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
 Data File : VU034483.D
 Acq On : 11 Sep 2019 18:42
 Operator : JC/SP
 Sample : PB122980TB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB122980TB

Quant Time: Sep 13 06:00:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	236587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	382841	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	387215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	211541	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.29	65	223677	62.49	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	124.98%	
35) Dibromofluoromethane	4.86	113	191630	53.79	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	107.58%	
50) Toluene-d8	7.54	98	703359	52.55	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	105.10%	
62) 4-Bromofluorobenzene	10.29	95	233020	45.09	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.33	43	11409	6.159	ug/l	95
18) Methyl Acetate	2.61	43	4885	1.084	ug/l	92
20) Methylene Chloride	2.69	84	5403	1.392	ug/l	93
43) Isopropyl Acetate	5.53	43	93754	12.243	ug/l	99
80) 1,3,5-Trimethylbenzene	10.75	105	13056	1.134	ug/l	94
95) Naphthalene	13.73	128	15123m	4.953	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
Data File : VU034483.D
Acq On : 11 Sep 2019 18:42
Operator : JC/SP
Sample : PB122980TB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB122980TB

Quant Time: Sep 13 06:00:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

