

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
 Data File : VU034494.D
 Acq On : 11 Sep 2019 22:57
 Operator : JC/SP
 Sample : K4624-23
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T-11-A3-(0)

Quant Time: Sep 12 05:58:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	248064	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	372561	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	399452	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	219233	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	215994	57.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.10%	
35) Dibromofluoromethane	4.86	113	187990	54.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.46%	
50) Toluene-d8	7.54	98	695313	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
62) 4-Bromofluorobenzene	10.29	95	235699	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
Target Compounds						
16) Acetone	2.31	43	15698	8.082	ug/l	96
20) Methylene Chloride	2.69	84	8302	2.228	ug/l	93
43) Isopropyl Acetate	5.53	43	49797	6.682	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
Data File : VU034494.D
Acq On : 11 Sep 2019 22:57
Operator : JC/SP
Sample : K4624-23
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T-11-A3-(0)

Quant Time: Sep 12 05:58:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

