

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
 Data File : VU034501.D
 Acq On : 12 Sep 2019 01:40
 Operator : JC/SP
 Sample : K4643-23
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T11-B2-(0)

Quant Time: Sep 12 06:38:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	236020	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	351919	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	377916	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	204356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	222118	62.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.40%	
35) Dibromofluoromethane	4.86	113	190278	58.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.22%	
50) Toluene-d8	7.54	98	673916	54.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.56%	
62) 4-Bromofluorobenzene	10.29	95	228062	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
Target Compounds						
16) Acetone	2.31	43	11533	6.241	ug/l	99
20) Methylene Chloride	2.69	84	14677	4.485	ug/l	97
43) Isopropyl Acetate	5.54	43	19931	2.831	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
Data File : VU034501.D
Acq On : 12 Sep 2019 01:40
Operator : JC/SP
Sample : K4643-23
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T11-B2-(0)

Quant Time: Sep 12 06:38:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

