

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
 Data File : VU034502.D
 Acq On : 12 Sep 2019 02:03
 Operator : JC/SP
 Sample : K4643-24
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T11-B3-(0)

Quant Time: Sep 12 06:41:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	241469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	362174	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	391820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	215076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	222123	60.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.60%	
35) Dibromofluoromethane	4.86	113	189168	56.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.26%	
50) Toluene-d8	7.54	98	701471	55.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.80%	
62) 4-Bromofluorobenzene	10.29	95	236859	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
Target Compounds						
16) Acetone	2.31	43	9636	5.097	ug/l	98
20) Methylene Chloride	2.69	84	10938	3.158	ug/l	96
43) Isopropyl Acetate	5.54	43	26441	3.650	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU091219\
Data File : VU034502.D
Acq On : 12 Sep 2019 02:03
Operator : JC/SP
Sample : K4643-24
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T11-B3-(0)

Quant Time: Sep 12 06:41:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

