

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091219\
 Data File : VU034507.D
 Acq On : 12 Sep 2019 09:56
 Operator : JC/SP
 Sample : VU0912WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0912WBL02

Quant Time: Sep 13 07:26:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	239704	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	347662	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	350209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	191498	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	201773	55.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.26%	
35) Dibromofluoromethane	4.86	113	177246	54.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.58%	
50) Toluene-d8	7.54	98	604949	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
62) 4-Bromofluorobenzene	10.29	95	203658	43.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091219\
Data File : VU034507.D
Acq On : 12 Sep 2019 09:56
Operator : JC/SP
Sample : VU0912WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0912WBL02

Quant Time: Sep 13 07:26:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

