

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091219\
 Data File : VU034508.D
 Acq On : 12 Sep 2019 10:23
 Operator : JC/SP
 Sample : K4756-25
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPA-RB1-M27W1-090419

Quant Time: Sep 13 07:28:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	213980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	303836	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	308698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	170382	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	183853	56.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.58%	
35) Dibromofluoromethane	4.86	113	162223	57.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.76%	
50) Toluene-d8	7.54	98	551151	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
62) 4-Bromofluorobenzene	10.29	95	189207	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.82	59	8980	13.129	ug/l #	89
16) Acetone	2.31	43	13148	7.848	ug/l	96
17) Carbon Disulfide	2.46	76	9282	1.774	ug/l #	94
20) Methylene Chloride	2.69	84	3127	0.746	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091219\
Data File : VU034508.D
Acq On : 12 Sep 2019 10:23
Operator : JC/SP
Sample : K4756-25
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPA-RB1-M27W1-090419

Quant Time: Sep 13 07:28:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

