

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091219\
 Data File : VU034526.D
 Acq On : 12 Sep 2019 17:37
 Operator : JC/SP
 Sample : K4756-19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPA-CPMW09S-M27W1-090319

Quant Time: Sep 13 07:59:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	260172	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.87	114	386839	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	395111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	230653	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	222694	56.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.14%	
35) Dibromofluoromethane	4.86	113	192217	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	
50) Toluene-d8	7.55	98	709642	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
62) 4-Bromofluorobenzene	10.29	95	242534	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.82	59	3060	3.679	ug/l #	90
16) Acetone	2.31	43	8843	4.341	ug/l	94
30) Chloroform	4.66	83	6305	0.984	ug/l	96
44) Trichloroethene	6.17	130	1136	0.303	ug/l	91
64) Tetrachloroethene	8.21	164	103222	30.113	ug/l	99
89) n-Butylbenzene	11.87	91	4132	0.347	ug/l	95
95) Naphthalene	13.74	128	15636	4.902	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091219\
Data File : VU034526.D
Acq On : 12 Sep 2019 17:37
Operator : JC/SP
Sample : K4756-19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPA-CPMW09S-M27W1-090319

Quant Time: Sep 13 07:59:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

