

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091219\
 Data File : VU034541.D
 Acq On : 12 Sep 2019 23:25
 Operator : JC/SP
 Sample : K4756-24RE
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPA-SW-18-M27W1-090619RE

Quant Time: Sep 13 08:23:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	230347	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	362811	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	363809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	205225	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	218690	62.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.50%	
35) Dibromofluoromethane	4.86	113	189908	56.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.50%	
50) Toluene-d8	7.54	98	642694	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
62) 4-Bromofluorobenzene	10.29	95	216199	44.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.30%	
Target Compounds						
16) Acetone	2.31	43	9460	5.245	ug/l	95
64) Tetrachloroethene	8.21	164	1515	0.480	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091219\
Data File : VU034541.D
Acq On : 12 Sep 2019 23:25
Operator : JC/SP
Sample : K4756-24RE
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPA-SW-18-M27W1-090619RE

Quant Time: Sep 13 08:23:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

