

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091219\
 Data File : VU034544.D
 Acq On : 13 Sep 2019 00:35
 Operator : JC/SP
 Sample : K4756-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPA-CPMW09D-M27W1-090319

Quant Time: Sep 13 08:28:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 06 05:10:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.96	168	238630	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.86	114	356394	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.07	117	370811	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.46	152	215055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.29	65	208234	57.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.34%	
35) Dibromofluoromethane	4.86	113	181015	54.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.16%	
50) Toluene-d8	7.55	98	648706	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
62) 4-Bromofluorobenzene	10.29	95	217503	45.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.42%	
Target Compounds						
16) Acetone	2.31	43	4389	2.349	ug/l	100
30) Chloroform	4.66	83	2765	0.471	ug/l #	83
64) Tetrachloroethene	8.21	164	93019	28.915	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091219\
Data File : VU034544.D
Acq On : 13 Sep 2019 00:35
Operator : JC/SP
Sample : K4756-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPA-CPMW09D-M27W1-090319

Quant Time: Sep 13 08:28:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U090619W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 06 05:10:13 2019
Response via : Initial Calibration

