

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091223\
 Data File : VU055260.D
 Acq On : 12 Sep 2023 20:08
 Operator : MD/SY
 Sample : VIBLK130
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK130

Quant Time: Sep 13 01:50:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 13 01:44:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	169909	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.419	117	187187	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	91227	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	33619	4.222	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.400%
7) Chloroethane-d5	1.904	69	51712	5.610	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.200%
11) 1,1-Dichloroethene-d2	2.564	65	17645	3.889	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.800%
20) 2-Butanone-d5	4.625	46	142698	61.750	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.500%
24) Chloroform-d	5.062	84	105828	5.008	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	5.698	65	58018	5.488	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.800%
32) Benzene-d6	5.724	84	199709	4.322	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
36) 1,2-Dichloropropane-d6	6.689	67	68967	4.693	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.898	98	174920	4.110	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	8.181	79	17091	4.223	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.400%
46) 2-Hexanone-d5	8.634	63	86438	52.608	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.220%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	52892	5.118	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.400%
66) 1,2-Dichlorobenzene-d4	12.193	152	63191	4.746	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.000%
Target Compounds						Qvalue
16) Methylene chloride	3.043	84	5753	0.479	ug/L	95
22) cis-1,2-Dichloroethene	4.660	96	1068	0.109	ug/L #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091223\
Data File : VU055260.D
Acq On : 12 Sep 2023 20:08
Operator : MD/SY
Sample : VIBLK130
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK130

Quant Time: Sep 13 01:50:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 13 01:44:08 2023
Response via : Initial Calibration

