

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091223\
 Data File : VU055267.D
 Acq On : 13 Sep 2023 03:01
 Operator : MD/SY
 Sample : 04350-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4M9

Quant Time: Sep 13 05:36:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 13 05:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	170154	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	180041	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	82211	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	36094	4.526	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.600%
7) Chloroethane-d5	1.907	69	41172	4.460	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.200%
11) 1,1-Dichloroethene-d2	2.563	65	14838	3.265	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.400%
20) 2-Butanone-d5	4.624	46	121073	52.317	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.640%
24) Chloroform-d	5.062	84	90114	4.258	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.200%
26) 1,2-Dichloroethane-d4	5.701	65	48237	4.556	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.200%
32) Benzene-d6	5.724	84	168041	3.781	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.600%
36) 1,2-Dichloropropane-d6	6.689	67	59755	4.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.600%
41) Toluene-d8	7.897	98	136191	3.327	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	66.600%#
43) trans-1,3-Dichloroprop...	8.180	79	12955	3.328	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.600%
46) 2-Hexanone-d5	8.634	63	63000	39.865	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.740%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	43585	4.384	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.600%
66) 1,2-Dichlorobenzene-d4	12.193	152	51291	4.274	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.400%
Target Compounds						
3) Chloromethane	1.515	50	1348	0.134	ug/L	96
5) Vinyl chloride	1.602	62	1435203	135.151	ug/L	95
12) 1,1-Dichloroethene	2.579	96	5699	0.724	ug/L #	4
17) Methyl tert-butyl Ether	3.361	73	27325	1.324	ug/L	97
18) trans-1,2-Dichloroethene	3.351	96	9628	1.206	ug/L	92
19) 1,1-Dichloroethane	3.869	63	15044	0.872	ug/L	94
22) cis-1,2-Dichloroethene	4.663	96	311771	31.912	ug/L	92
33) Benzene	5.772	78	4721	0.123	ug/L	100
34) Trichloroethene	6.541	95	46768	4.392	ug/L	98
42) Toluene	7.968	91	19045	0.466	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091223\
Data File : VU055267.D
Acq On : 13 Sep 2023 03:01
Operator : MD/SY
Sample : 04350-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4M9

Quant Time: Sep 13 05:36:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 13 05:34:47 2023
Response via : Initial Calibration

