

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091223\
 Data File : VU055270.D
 Acq On : 13 Sep 2023 04:15
 Operator : MD/SY
 Sample : 04350-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4N2

Quant Time: Sep 13 05:37:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 13 05:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	178848	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	187968	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	84008	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	26799	3.197	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.000%
7) Chloroethane-d5	1.907	69	40811	4.206	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.200%
11) 1,1-Dichloroethene-d2	2.563	65	15121	3.166	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.400%
20) 2-Butanone-d5	4.624	46	135483	55.698	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.400%
24) Chloroform-d	5.062	84	99188	4.459	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	5.698	65	52047	4.677	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.724	84	182667	3.937	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
36) 1,2-Dichloropropane-d6	6.689	67	61621	4.176	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.600%
41) Toluene-d8	7.898	98	158383	3.706	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.200%
43) trans-1,3-Dichloroprop...	8.181	79	13938	3.430	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.600%
46) 2-Hexanone-d5	8.631	63	76213	46.192	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.380%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	47382	4.565	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.400%
66) 1,2-Dichlorobenzene-d4	12.193	152	52809	4.307	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.200%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.383	85	2723	0.262	ug/L #	85
3) Chloromethane	1.518	50	1478	0.140	ug/L	87
5) Vinyl chloride	1.602	62	114728	10.279	ug/L	98
12) 1,1-Dichloroethene	2.576	96	170646	20.638	ug/L #	75
17) Methyl tert-butyl Ether	3.361	73	94966	4.379	ug/L	99
18) trans-1,2-Dichloroethene	3.354	96	7339	0.875	ug/L	99
19) 1,1-Dichloroethane	3.866	63	72507	3.999	ug/L	97
22) cis-1,2-Dichloroethene	4.660	96	6939633	675.786	ug/L	92
25) Chloroform	5.084	83	55633	2.946	ug/L	92
27) 1,2-Dichloroethane	5.795	62	18313	1.589	ug/L	96
29) 1,1,1-Trichloroethane	5.316	97	16025	0.986	ug/L #	88
33) Benzene	5.772	78	5824	0.146	ug/L	100
34) Trichloroethene	6.541	95	2879494	258.990	ug/L	98
47) Tetrachloroethene	8.553	164	23409	3.046	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091223\
Data File : VU055270.D
Acq On : 13 Sep 2023 04:15
Operator : MD/SY
Sample : 04350-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4N2

Quant Time: Sep 13 05:37:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 13 05:34:47 2023
Response via : Initial Calibration

