

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091224\
 Data File : VU060718.D
 Acq On : 12 Sep 2024 16:20
 Operator : MD/SY
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5101

Quant Time: Sep 13 05:47:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 05:03:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	4624	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	7108	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	2209	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	3921	0.521	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 104.000%			
4) 1,2-Dichloroethane-d4	5.709	65	2053	0.515	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 104.000%			
7) 1,2-Dichloropropane-d6	6.687	67	1774	0.401	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 80.000%			
8) Toluene-d8	7.887	98	4397	0.482	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 96.000%			
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1466	0.379	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 76.000%			
Target Compounds						Qvalue
3) Vinyl chloride	1.597	62	4428	0.561	ug/L	99
6) Trichloroethene	6.544	95	1878	0.418	ug/L	97
9) 1,2-Dibromoethane	8.924	107	1245	0.442	ug/L #	98
12) 1,2,3-Trichloropropane	10.814	75	1118	0.436	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.990	75	248	0.498	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091224\
Data File : VU060718.D
Acq On : 12 Sep 2024 16:20
Operator : MD/SY
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5101

Quant Time: Sep 13 05:47:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 05:03:57 2024
Response via : Initial Calibration

